

Instruction Guide



FluorPen FP 110
PAR-FluorPen FP 110
Monitoring Pen MP 100

Please read the Guide before operating this product



Manual Version: 2021/12

© PSI (Photon Systems Instruments), spol. s r.o.

www.psi.cz

This document and its parts can be copied or provided to a third party only with the express permission of PSI.

The contents of this manual have been verified to correspond to the specifications of the device. However, deviations cannot be ruled out. Therefore, a complete correspondence between the manual and the real device cannot be guaranteed. The information in this manual is regularly checked, and corrections may be made in subsequent versions.

The visualizations shown in this manual are only illustrative.

This manual is an integral part of the purchase and delivery of equipment and its accessories and both Parties must abide by it.


TABLE OF CONTENT

1	Information before using FluorPen device	5
2	General Description	6
2.1	<i>Technical Specification</i>	8
3	Device Description	11
3.1	<i>List of equipment and customer information</i>	12
3.2	<i>Care and maintenance</i>	12
4	Principle of measurement	13
5	Getting started	15
5.1	<i>Pulses description and setting</i>	15
5.2	<i>Measurement</i>	18
5.3	<i>Protocols explanation</i>	18
5.3.1	<i>OJIP protocol</i>	18
5.3.2	<i>Non-photochemical quenching (NPQ) protocols</i>	20
5.3.3	<i>Light curve (LC) protocols</i>	23
5.4	<i>Multiple measurement</i>	26
6	Control menu tree	27
7	USB Connection	35
8	Bluetooth connection	36
8.1	<i>Bluetooth pairing</i>	36
9	FluorPen software	39
9.1	<i>Software installation</i>	39
9.2	<i>Menu and icons explanation</i>	40
9.2.1	<i>Main menu</i>	40
9.2.2	<i>Menu settings</i>	41
9.2.3	<i>Menu online control</i>	41
9.3	<i>Data transfer and visualization</i>	44
9.4	<i>Firmware update</i>	46
10	GPS module	48
10.1	<i>GPS/FluorPen operation</i>	48
10.2	<i>Data download</i>	48
11	Warranty terms and conditions	50

12	Troubleshooting and Customer Support	50
13	Appendix	51
13.1	<i>Battery pack for Monitoring Pen</i>	51
13.1.1	Standard battery pack	51
13.1.2	Extended temperature range battery pack	52
13.2	<i>Installation and operation of the Monitoring Pen MP 100-A/B</i>	53
13.2.1	Device control	53
13.2.2	Device installation	55
13.3	<i>Detachable leaf clips</i>	55

1 INFORMATION BEFORE USING FLUORPEN DEVICE

Read this manual carefully before operating the device. If you are not sure about something in the manual, contact the manufacturer for clarification.

	By accepting the device, the customer agrees to follow the instructions in this guide.
---	--

Always follow corresponding manuals while working with the FluorPen device or doing the maintenance.

It is forbidden to interfere with the hardware or software of the FluorPen device in any way without previous agreement with the manufacturer.

The following table presents basic highlight symbols used in this manual:



Symbol	Description
	Important information, read carefully.
	Complementary and additional information.

Table 1 Used symbols.

2 GENERAL DESCRIPTION

FluorPen FP 110 is a portable, battery-powered fluorometer that enables quick and precise measurement of chlorophyll fluorescence parameters (Ft, QY, NPQ, OJIP, and LC of (QY) in plants. The FluorPen can be used in the laboratory, greenhouse, or in the field, where data can be mapped precisely to the location with the built in GPS module. With the built-in GPS module, the FluorPen is a great device for studying photosynthetic activity, stress detection, herbicide testing, or mutant/variety/transgenic screening in the field. Affordable price and user-friendly two-button operation make the FluorPen a perfect tool for teaching photosynthesis. Because of its rapid measurement capability and large internal memory, the FluorPen is also an invaluable tool for large plant-screening programs.

PAR-FluorPen FP 110 includes all features of the FluorPen FP 110, and measures the same chlorophyll fluorescence parameters as FP110, but in addition, the PAR-FluorPen has an integrated Light Meter for direct digital readouts of Photosynthetically Active Radiation (PAR) in the range from 400 to 700 nm. PAR is measured as Photosynthetic Photon Flux Density (PPFD), which is indicated by units of quanta (photons) per unit time per unit surface area. The sensor has a uniform response to photons in the 400-700 range. Instant readouts are provided as average values of 20 measurements. It is recommended to recalibrate the PAR sensor every 2 years.

Monitoring Pen MP 100 is a lightweight, portable and a more durable version of the FluorPen. It is designed for long-term, unattended monitoring of chlorophyll fluorescence parameters in the field or lab experiments. It features weatherproof construction for use even in adverse environmental conditions, or in the laboratory /greenhouse long term experiments. It is battery operated (internal or external battery) and as an option may be used with a solar panel as a power supply.

All measured data are sequentially stored in the internal memory of the FluorPen, PAR-FluorPen or Monitoring Pen all collected data can be transferred from the devices to the PC computer via both USB and Bluetooth communication -. Comprehensive FluorPen 1.1 software, included with the device provides data transfer, and visualization protocols.



Unless stated otherwise, the information regarding the FluorPen FP 110 is relevant also to PAR FluorpenFP110 and the Monitoring Pen MP 100.

FluorPen versions:

FluorPen FP 110/S	Equipped with a standard attached leaf-clip.
FluorPen FP 110/D	Adapted for use with detachable leaf-clips; leaf clips sold separately in sets of 10.
FluorPen FP 110/P (Fig. 1c)	Intended for autonomous use in indoor conditions (previously Monitoring Pen-S). It features a plastic case, measuring probe and thread for tripod attachment.
FluorPen FP 110/X	The "X" version is mounted with custom-made leaf-clip.
PAR-FluorPen FP 110/S (Fig. 1a)	Same features as the standard FluorPen FP 110/S plus Photosynthetically Active Radiation (PAR) meter in the range from 400 to 700 nm. equipped with a standard leaf-clip.
PAR-FluorPen FP 110/D (Fig. 1b)	Same features as the FluorPen FP 110/D adapted for use with detachable leaf clips plus Photosynthetically Active Radiation (PAR) meter in the range from 400 to 700 nm. Leaf clips sold separately.
PAR-FluorPen FP 110/X	Same features as the PAR-FluorPen FP 110/D but the "X" version is mounted with custom-made leaf-clip.




Fig. 1 a) PAR-FluorPen FP 110/S. b) PAR-FluorPen FP 110/D. c) FluorPen FP 110/P.

Monitoring Pen versions:

Monitoring Pen MP 100-E (Fig. 2a)

Monitoring Pen MP 100-E is a modified FluorPen designed for extra durability, battery-powered and intended for autonomous use in field conditions during extended experiments. It features waterproof metal case, measuring probe, thread for tripod attachment, external pack with batteries and the FluorPen 1.1 software for data collection and processing.

Monitoring Pen MP 100-A (Fig. 2b)

This is a submersible, battery-powered FluorPen intended for underwater measurements of chlorophyll fluorescence parameters (also autonomous). It features a waterproof case, measuring probe, and the FluorPen 1.1 software for data collection and processing. External battery pack with batteries is sold separately. This aquatic version of the monitoring pen is intended for use at maximum water depth of 2 meters. The device is equipped with two buttons that allow direct control of the device (even under water). A customized Version B of this instrument is also available for use in deeper water (maximum 10 m). There are no control buttons on this version of the device. Version B is controlled via software and a PC (placed above water). The device is usually fixed in static position under water.



Fig. 2 a) Monitoring Pen MP 100-E. b) Monitoring Pen MP 100-A.

2.1 TECHNICAL SPECIFICATION

FluorPen and PAR-FluorPen

Protocols	Ft – instantaneous chlorophyll fluorescence Quantum Yield OJIP Non-photochemical quenching Light curve Photosynthetically Active Radiation (measured as PPFD) – PAR-FluorPen only
LED emitter	Blue (470 nm), other wavelengths on request
Saturating pulse illumination	Up to 3,000 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ (adjustable from 10 to 100%)
Actinic illumination	Adjustable from 10 to 1,000 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$
Measuring illumination	Up to 0,09 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ per pulse (adjustable from 10 to 100%)
Detector	PIN photodiode with bandpass filters Wavelength range from 667 to 750 nm
Internal memory capacity	Up to 16 Mb
Internal data logging	Up to 149,000 measurements (depending on protocol)
Data transfer	USB cable Bluetooth (transfer up to 3Mbps for distance up to 20m)
PC software	FluorPen 1.1 (Windows 7 and higher)
Battery	Li-Ion rechargeable battery Capacity 2000 mAh Max. charging current 0.5 A Charging via USB port - PC, power bank, USB charger, etc. 48 hours typical with full operation Low battery indicator
Sample holder	Standard leaf-clip (FP 110/S) Detachable leaf-clip (FP 110/D) Probe (FP 110/P)
Display	Graphical display
Keypad	Sealed, 2-key tactile response Turns off after 5 minutes of no use
Built in GPS module	Ultra-high sensitivity down to -165dBm High accuracy of <1.5 m in 50% of trials
Size	134 x 65 x 33 mm
Weight	188 g
Operating conditions	Temperature: 0 to +55 °C Relative humidity: 0 to 95 % (non-condensing)
Storage conditions	Temperature: -10 to +60 °C Relative humidity: 0 to 95 % (non-condensing)
Warranty	1-year parts and labor

Monitoring Pen

Protocols	Ft – instantaneous chlorophyll fluorescence Quantum Yield OJIP Non-photochemical quenching Light curve	
LED emitter	Blue (470 nm), other wavelengths on request	
Saturating pulse illumination	Up to 3,000 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ (adjustable from 10 to 100%)	
Actinic illumination	Adjustable from 10 to 1,000 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$	
Measuring illumination	Up to 0,09 $\mu\text{mol}(\text{photon})\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ per pulse (adjustable from 10 to 100%)	
Detector	PIN photodiode with bandpass filters Wavelength range from 667 to 750 nm	
Internal memory capacity	Up to 16 Mb	
Internal data logging	Up to 149,000 measurements (depending on protocol)	
Data transfer	Serial cable	
PC software	FluorPen 1.1 (Windows 7 and higher)	
External Battery	Standard battery pack operating temperature from 10 to 40 °C rechargeable	Extended temperature range battery pack operating temperature from -40 to 60 °C non-rechargeable (spare battery)
	Capacity 12Ah Battery life up to 2 years of operation (1 QY measurement per hour)	
Sample holder	Probe	
Display	2 x 8 characters LC display	
Keypad	Sealed, 2-key tactile response Turns off after 5 minutes of no use	
Built in GPS module	Ultra-high sensitivity down to -165dBm High accuracy of <1.5 m in 50% of trials	
Size	134 x 65 x 33 mm	
Weight	188 g	
Warranty	1-year parts and labor	

Bluetooth module compliance data:

Category	Country	Standard
Radio	USA	FCC Part 15 Subpart B: 2008 Class B FCC CRF Title 47 Part 15 Subpart C
	FCC ID:	T9J-RN42
	Europe	ETSI EN 301 489-1 V1.8.1 ETSI EN 301 489-17 V2.1.1 ETSI EN 300 328 V1.7.1
	Canada	IC RSS-210 low power comm. device
	Certification number:	6514A-RN42
EMC	USA	FCC CFR47 Part 15 subclass B
	Europe	EN 55022 Class B radiated EN61000-4-2 ESD immunity EN61000-4-3 radiated field EN61000-4-6 RF immunity EN61000-4-8 power magnetic immunity

3 DEVICE DESCRIPTION




Fig. 3 Device description.

3.1 LIST OF EQUIPMENT AND CUSTOMER INFORMATION

Carefully unpack the carton. You should have received the following items:

- FluorPen/Monitoring Pen
- Carrying Case
- Textile Strap for Comfortable Wearing
- FluorPen Operating Manual (on a USB flash disc)
- FluorPen software and driver (on a USB flash disc)
- USB cable
- Self-Adhesive Rubber Pads for Optics Protection (FP 110/S only)
- Detachable Leaf-clips (FP 110/D only and sold separately)

Other Accessories or Optional Features (according to your specific order)

	<p>For data download via USB connection, the USB driver needs to be installed on the PC. It can be found on the installation disk (USB driver folder).</p> <p>If any item is missing, please, contact the manufacturer. Also check the carton for any visible external damage. If any damage is found, notify the carrier and the manufacturer immediately. The carton and all packing materials should be retained for inspection by the carrier or insurer.</p> <p>For customer support, please write to: support@psi.cz</p>
--	--

3.2 CARE AND MAINTENANCE

FluorPen and Monitoring Pen

- Never submerge the device in water! (except Monitoring Pen MP 100-A).
- The device should not come in contact with any organic solvents, strong acids or bases.
- Keep the optical part clean and dry. If cleaning is needed, use soft, non-abrasive tissue.
- Battery charge lasts approximately 48 hours when the FluorPen is operated continuously.
- If the battery can no longer be charged please contact PSI for replacement battery and installation instructions.

Li-ion battery

- Avoid fully discharging of the battery.
- Do not keep the battery at full charge for all the time.
- Keeping at high temperatures shortens battery life.

4 PRINCIPLE OF MEASUREMENT

FluorPen is a chlorophyll fluorometer and is used to measure different photosynthetic parameters in plants. It is equipped with a **blue LED emitter (470 nm)**, optically filtered and precisely focused to deliver light intensities of up to $3,000 \mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ to measured plant tissue (Fig. 4).



Fig. 4 Fluor Pen FP 110/S.

When studying photosynthesis using chlorophyll fluorescence, researchers must distinguish between **photochemical quenching** and **non-photochemical quenching** (heat dissipation). This is achieved by stopping photochemistry, which allows researchers to measure fluorescence in the presence of non-photochemical quenching alone. To reduce photochemical quenching to negligible levels, a high intensity, short flash of light is applied to the leaf. This transiently closes all PSII reaction centers, which prevents energy of PSII being passed to downstream electron carriers. Non-photochemical quenching will not be affected if the flash is short. During the flash, the fluorescence reaches the level obtained in the absence of any photochemical quenching, known as **maximum fluorescence F_m** . The efficiency of photochemical quenching (which is a proxy of the efficiency of PSII) can be estimated by comparing F_m to the **steady yield of fluorescence in the light F_t** and the yield of fluorescence in the **absence of photosynthetic light F_0** . The efficiency of non-photochemical quenching is altered by various internal and external factors. Alterations in heat dissipation result in changes in F_m . Heat dissipation cannot be totally stopped, so the yield of chlorophyll fluorescence in the absence of non-photochemical quenching cannot be measured. See picture below (Fig. 5). When measuring F_m it is important to dark adapt the samples. This can be achieved by placing the sample in the dark for few minutes (the time varies with conditions) or by using the FP110/D version of the FluorPen that has been adapted for detachable leaf clips. The leaf clips may be placed on the leaf ahead of the measurements and once dark adaptation has been achieved the FP-110/D may be attached to the leaf clip without exposing the leaf to light.




Fig. 5 Chlorophyll fluorescence.

FluorPen measures:

<p>Ft - Instantaneous Chlorophyll Fluorescence F_t is equivalent to F_0 if the sample is dark-adapted.</p>
<p>QY - Quantum Yield QY is a measure of the Photosystem II efficiency. QY is equivalent to F_v/F_m in dark-adapted samples and to F_v'/F_m' in light-adapted samples.</p>
<p>OJIP - Chlorophyll Fluorescence Induction Kinetics The OJIP curves show major changes that occur during exposure of a sample to high irradiance (see more in chapter 5.3.1).</p>
<p>NPQ - Non-Photochemical Quenching The NPQ protocol is used to quantify photochemical and non-photochemical quenching. The measurement should be performed with a dark-adapted sample. (see more in chapter 5.3.2).</p>
<p>LC - Light Curve Photosystem II Quantum Yield estimated from fluorescence that is measured sequentially at several different light levels. More in chapter 5.3.3.</p>
<p>PAR - Photosynthetically Active Radiation (PAR-FluorPen only) Photosynthetically Active Radiation measured as Photosynthetic Photon Flux Density (PPFD).</p>

5 GETTING STARTED

For more detailed information on particular steps of FluorPen operation please refer to chapter 6.

The device can be powered with built-in Li-Ion battery. Ensure the battery is fully charged by plugging it into a PC via USB cable or the AC outlet via the USB cable and a USB adaptor (not included). Monitoring Pen can be powered from an optional battery pack (see more in chapter 13.1).

The FluorPen is controlled using two buttons:

- Use the **MENU** key to scroll through sequential menu options on the digital display and to turn the device off (hold for 3 s).
- Use the **SET** key to turn the device on (hold for 1 sec) and select a menu option based on cursor (>) position.

5.1 PULSES DESCRIPTION AND SETTING

Flash pulse

This function serves for setting of measuring pulses intensity. The measuring pulses are weak light pulses, which are able to induce the minimal chlorophyll fluorescence (F_0 or F_t). It takes only 30 μs and the maximum intensity is 3,000 $\mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$. It means $30 \mu\text{s} * 3,000 \mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1} = 0.09 \mu\text{mol}\cdot\text{m}^{-2}$ per pulse is the maximal intensity of the flash pulse.

Super pulse

This function serves for setting intensity of the saturating pulse. Saturating light pulse is able to induce maximum chlorophyll fluorescence (F_m). 100% of intensity represents approximately 3,000 $\mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$.

Actinic pulse


This function serves for setting intensity of actinic light. It is the ambient light in which the plants are growing. 100% of intensity equals approximately 1,000 $\mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$.

Pulses used in predefined protocols:

Measurements based on fluorescence	Used pulses
F_t	Flash pulse
QY	Flash pulse, Super pulse
OJIP	Super pulse
NPQ protocols	Flash pulse, Super pulse, Actinic pulse
Light Curves	Flash pulse, Super pulse (Actinic pulse is preset)

Default setting of light color and intensities in FluorPen firmware. These may be changed according to user requirements and samples growth conditions:

Flash pulse 30% = Measuring flash pulse
Super pulse 80% = Saturating pulse
Actinic pulse 300 $\mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$ (30 %) = Actinic light

	Please note that those parameters are recommended by manufacturer and can be change according to user needs.
---	--

Setting of optimal intensities of pulses:

Flash pulse setting

The optimum value of Flash pulse can be determined with QY measurement as shown in Fig.6 below. Before performing QY measurement it is recommended to set the intensity of Super pulse to 80 %.

Please note that QY measurement should be performed with dark adapted sample, therefore the same sample (position on the leaf) should not be used more than once unless dark adaptation follows the first measurement. The recommendation is to use a new sample (new area on the leaf) for each QY measurements.

F_0 increases linearly with growing intensity of the Flash pulse.

The Flash pulse setting recommended by manufacturer is 30 %. One can increase the intensity of Flash pulse for samples with very low chlorophyll density. However, it should be noted that high intensities of Flash pulse can cause undesirable “actinic effect” as higher intensity High Flash pulse will initiate the photochemistry. Changes in the Flash pulse will affect F_0 and the QY value will be lower.

The optimal Flash pulse intensity is that at which the highest value of QY is reached. This can be easily determined on one leaf by measuring QY in few different spots with different flash pulse settings. See (Fig. 6) below. In this example the optimal flash pulse setting is 30%.

595		596		597	
15:17:42 19.7.2016		15:19:01 19.7.2016		15:20:03 19.7.2016	
QY		QY		QY	
0.71		0.69		0.68	
Fo Backgr	289	Fo Backgr	289	Fo Backgr	390
Fo Flash	2552	Fo Flash	4426	Fo Flash	8875
Fm Backgr	309	Fm Backgr	269	Fm Backgr	390
Fm Flash	7995	Fm Flash	13419	Fm Flash	26659
30% f_pulse		50% f_pulse		100 % f_pulse	

Fig. 6 QY measurement performed with different intensities of Flash pulse. Optimal setting is highlighted in red rectangle.

Super pulse setting

To determine the optimal intensity of Super pulse is to perform OJIP measurement with different Super pulse settings.

Please note that OJIP measurement should be performed with dark adapted sample. New sample (new section of the same leaf) should be used for every measurement as exposure to super pulse will change photochemistry of the leaf in that section.

The Super pulse setting recommended by manufacturer is 80 %.

When performing the OJIP measurement with different intensities of Super pulse the Fv/Fm value will stop increasing when the optimal level has been reached for the samples used (Fig. 7 and Fig. 8).




Fig. 7 OJIP measurement performed with different intensities of Super pulse.

80	82	83	84	85
8:42:40 22.1.2019	8:57:55 22.1.2019	9:00:08 22.1.2019	9:02:07 22.1.2019	9:04:29 22.1.2019
OJIP	OJIP	OJIP	OJIP	OJIP
Bckg 357	Bckg 390	Bckg 390	Bckg 390	Bckg 390
Fo 6405	Fo 12451	Fo 19019	Fo 26659	Fo 36184
Fj 10956	Fj 27276	Fj 45905	Fj 64664	Fj 89275
Fi 19637	Fi 43987	Fi 70581	Fi 98736	Fi 132742
Fm 28220	Fm 53805	Fm 81115	Fm 110082	Fm 132742
Fv 21815	Fv 41354	Fv 62096	Fv 83423	Fv 96558
Vj 0.209	Vj 0.358	Vj 0.433	Vj 0.456	Vj 0.550
Vi 0.607	Vi 0.763	Vi 0.830	Vi 0.864	Vi 1.000
Fm/Fo 4.406	Fm/Fo 4.321	Fm/Fo 4.265	Fm/Fo 4.129	Fm/Fo 3.669
Fv/Fo 3.406	Fv/Fo 3.321	Fv/Fo 3.265	Fv/Fo 3.129	Fv/Fo 2.669
Fv/Fm 0.773	Fv/Fm 0.769	Fv/Fm 0.766	Fv/Fm 0.758	Fv/Fm 0.727
Mo 0.155	Mo 0.343	Mo 0.547	Mo 0.728	Mo 1.083
Area 10680744	Area 15155293	Area 19219604	Area 28544434	Area 12998559
Fix Area 26970048	Fix Area 51974884	Fix Area 78608144	Fix Area 107026000	Fix Area 132466992
HACH Are 20565304	HACH Are 39524384	HACH Are 59589908	HACH Are 80368064	HACH Are 96284440
Sm 489.605	Sm 366.477	Sm 309.514	Sm 342.165	Sm 134.619
Ss 1.346	Ss 1.046	Ss 0.792	Ss 0.626	Ss 0.508
N 363.627	N 350.434	N 390.721	N 546.744	N 265.115
Phi_Po 0.773	Phi_Po 0.769	Phi_Po 0.766	Phi_Po 0.758	Phi_Po 0.727
Psi_o 0.791	Psi_o 0.642	Psi_o 0.567	Psi_o 0.544	Psi_o 0.450
Phi_Eo 0.612	Phi_Eo 0.493	Phi_Eo 0.434	Phi_Eo 0.413	Phi_Eo 0.327
Phi_Do 0.227	Phi_Do 0.231	Phi_Do 0.234	Phi_Do 0.242	Phi_Do 0.273
Phi_Pav 919.490	Phi_Pav 933.856	Phi_Pav 941.665	Phi_Pav 954.645	Phi_Pav 963.615
Pi_Abs 13.448	Pi_Abs 4.777	Pi_Abs 2.593	Pi_Abs 1.774	Pi_Abs 0.807
ABS/RC 0.961	ABS/RC 1.244	ABS/RC 1.649	ABS/RC 2.109	ABS/RC 2.707
TRo/RC 0.743	TRo/RC 0.956	TRo/RC 1.262	TRo/RC 1.598	TRo/RC 1.969
ETo/RC 0.588	ETo/RC 0.613	ETo/RC 0.716	ETo/RC 0.870	ETo/RC 0.887
Dio/RC 0.218	Dio/RC 0.288	Dio/RC 0.387	Dio/RC 0.511	Dio/RC 0.738
FLASH	FLASH	FLASH	FLASH	FLASH
[nm] 455	[nm] 455	[nm] 455	[nm] 455	[nm] 455
[%] 30	[%] 30	[%] 30	[%] 30	[%] 30
[uE] -NAN	[uE] -NAN	[uE] -NAN	[uE] -NAN	[uE] -NAN
SUPER	SUPER	SUPER	SUPER	SUPER
[nm] 455	[nm] 455	[nm] 455	[nm] 455	[nm] 455
[%] 20	[%] 40	[%] 60	[%] 80	[%] 100
[uE] -NAN	[uE] -NAN	[uE] -NAN	[uE] -NAN	[uE] -NAN
...
20% F_pulse	40% F_pulse	60% F_pulse	80% F_pulse	100% F_pulse

Fig. 8 OJIP data - measurement performed with different intensities of Super pulse. The highest Fv/Fm value indicates the optimal intensity of Super pulse (20% in this case).

Actinic pulse setting

Intensity of Actinic pulse should correspond with cultivation light intensity or should be set according to application.



Should **Overflow** be observed on the display during the measurement, lower the intensity of the used pulses.
In case of **Low value** on display during measurement, increase the intensity of the used pulses.

5.2 MEASUREMENT

No device calibration is needed before chlorophyll fluorescence measurements are made. Results of fluorescence measurement depend on device settings and the samples.

How to perform Chlorophyll Fluorescence measurement with FluorPen:

- Prepare dark adapted sample first (prior to F0, QY, NPQ, LC measurements) by placing the sample for at least 10-15 min in the dark. Alternatively, dark adaptation can be easily achieved by placing the detached leaf clips in closed position on the leaf ahead of the measurements. Only the FP110/D or PAR-FP110/D is designed for use with the detachable leaf clips. The duration of dark-adaptation period depends on plant species and growth conditions.
- For light adapted measurements no dark adaptation of the sample is required.
- Turn ON the device by holding the **SET** button for 1 sec.
- Place the dark-adapted leaf in the leaf-clip (FP110/S or PAR-FP110/S) or in case of detachable leaf clips place the leaf clip on the optical probe of the FluorPen and slide open the screen of the leaf clip to expose the leaf to the optical probe.
- Select **Measure > from the menu and select** required parameter for example – **QY** (press SET as Enter button when making selections).
- Press **SET** to start the measurements.
- When **OJIP**, **LC** or **NPQ** are being measured the display on the device shows the progress of the measurement as percentage.
- When Ft or QY are measured the values appear on the device display. The result of OJIP, NPQ or LC protocol are not visible on the display of the device and need to be download to PC computer (via USB cable or BT connection, see instructions in Chapter 7 and 8) using FluorPen Software (downloaded to PC earlier).
- All measured data are stored in the device memory and can be downloaded to PC computer after completion of the experiment.

5.3 PROTOCOLS EXPLANATION

5.3.1 OJIP PROTOCOL

The FluorPen device offers the protocol to capture rapid fluorescence transient – OJIP, which occurs during exposure of photosynthetic organisms to high irradiance. The FluorPen software enables data downloading to a PC and subsequent OJIP curve and calculated data visualization.

The OJIP protocol includes the following measured and calculated parameters:

Abbreviation	Explanation
Bckg	Background
F_0	$F_0 = F_{50\mu s}$, fluorescence intensity at 50 μs
F_j	F_j = fluorescence intensity at J-step (at 2 ms)
F_i	F_i = fluorescence intensity at i-step (at 30 ms)
F_m	F_m = maximal fluorescence intensity
F_v	$F_v = F_m - F_0$ (maximal variable fluorescence)
V_j	$V_j = (F_j - F_0) / (F_m - F_0)$
V_i	$V_i = (F_i - F_0) / (F_m - F_0)$
F_m / F_0	
F_v / F_0	
F_v / F_m	
M_0 or $(dV/dt)_0$	$M_0 = TR_0 / RC - ET_0 / RC = 4 (F_{300} - F_0) / (F_m - F_0)$
Area	Area between fluorescence curve and F_m (background subtracted)
Fix Area	Area below the fluorescence curve between $F_{40\mu s}$ and F_{1s} (background subtracted)
S_M	$S_M = Area / (F_m - F_0)$ (multiple turn-over)
S_S	S_S = the smallest S_M turn-over (single turn-over)
N	$N = S_M \cdot M_0 \cdot (1 / V_j)$ turn-over number Q_A
Φ_{P_0}	$\Phi_{P_0} = 1 - (F_0 / F_m)$ (or F_v / F_m)
Ψ_{i_0}	$\Psi_{i_0} = 1 - V_j$
Φ_{E_0}	$\Phi_{E_0} = (1 - (F_0 / F_m)) \cdot \Psi_{i_0}$
Φ_{D_0}	$\Phi_{D_0} = 1 - \Phi_{P_0} = (F_0 / F_m)$
$\Phi_{P_{av}}$	$\Phi_{P_{av}} = \Phi_{P_0} (S_M / t_{Fm})$ t_{Fm} = time to reach F_m (in ms)
ABS / RC	$ABS / RC = M_0 \cdot (1 / V_j) \cdot (1 / \Phi_{P_0})$
TR_0 / RC	$TR_0 / RC = M_0 \cdot (1 / V_j)$
ET_0 / RC	$ET_0 / RC = M_0 \cdot (1 / V_j) \cdot \Psi_{i_0}$
DI_0 / RC	$DI_0 / RC = (ABS / RC) - (TR_0 / RC)$

Formulas Derived From:

R.J. Strasser, A. Srivastava and M. Tsimilli-Michael (2000): The fluorescence transient as a tool to characterize and screen photosynthetic samples. In: Probing Photosynthesis: Mechanism, Regulation and Adaptation (M. Yunus, U. Pathre and P. Mohanty, eds.), Taylor and Francis, UK, Chapter 25, pp 445-483.

5.3.2 NON-PHOTOCHEMICAL QUENCHING (NPQ) PROTOCOLS

The NPQ protocol is used to quantify photochemical and non-photochemical quenching. It should be performed with dark-adapted samples. The NPQ protocol starts with measurement of minimal level of fluorescence F_0 during a dark period. A short saturating flash of light is then applied to reduce the plastoquinone pool and measure maximum fluorescence in the dark-adapted state, F_m . After a short dark relaxation, the sample is exposed to actinic irradiance for tens to hundreds of seconds to elicit a transient called the Kautsky effect. A sequence of saturating flashes is then applied during exposure to actinic light to probe the non-photochemical quenching NPQ and effective quantum yield of photosynthesis QY in light adapted state. After exposure to continuous illumination, the relaxation of non-photochemical quenching is determined by means of saturating pulses applied in dark. This sequence of the protocol is illustrated in Fig. 9.

The FluorPen device comes with three predefined NPQ protocols, NPQ1, NPQ2 and NPQ3. The protocols differ in the duration of the light exposure and the dark recovery phase, and in the number and interval between pulses. See Table 2:

	Phase	Duration	# of pulses	1st pulse	Pulse interval
NPQ1	Light	60 s	5	7 s	12 s
	Dark recovery	88 s	3	11 s	26 s
NPQ2	Light	200 s	10	10 s	20 s
	Dark recovery	390 s	7	20 s	60 s
NPQ3	Light	200 s	10	11 s	21 s
	Dark recovery	60 s	2	20 s	21 s

Table 2 NPQ Protocols.

The NPQ protocols include the following measured and calculated parameters:

Abbreviation	Explanation
F_0	minimum fluorescence in dark-adapted state
F_m	maximum fluorescence in dark-adapted state, measured during the first saturation flash after dark adaptation
F_p	fluorescence in the peak of fast Kautsky induction
F_{m_Ln} , Lss, D, Dn ¹	maximum fluorescence
QY_{max}^2	maximum quantum yield of PSII in dark-adapted state - F_v/F_m
QY_Ln , Lss, D, Dn ^{1,3}	effective quantum yield of PSII
NPQ_Ln , Lss, D, Dn ^{1,4}	non-photochemical chlorophyll fluorescence quenching
Qp_Ln , Lss, D, Dn ^{1,5}	coefficient of photochemical quenching, an estimate of open PSII reaction centers

¹ L - indicates light adapted parameters; D - refers to dark recovery phase after switching of the actinic illumination; n - represents a sequential number of light phases; ss - steady state

² Calculated as $(F_m - F_0) / F_m$

³ Calculated as $(F_{m_Ln} - F_{t_Ln}) / F_{m_Ln}$ or of corresponding steady state or dark recovery parameters

⁴ Calculated as $(F_m - F_{m_Ln}) / F_{m_Ln}$ or of corresponding ss, Dn or Dss parameters

⁵ Calculated as $(F_{m_Ln} - F_{t_Ln}) / (F_{m_Ln} - F_{0_Ln})$ or of corresponding ss, Dn or Dss parameters

F_0 _Ln is calculated as $F_0 / ((F_m - F_0) / F_m + F_0 / F_{m_Ln})$.

For more details, please refer to: Oxborough K., Baker N.R. (1997): Resolving chlorophyll a fluorescence images of photosynthetic efficiency into photochemical and non-photochemical components: calculation of qP and Fv'/Fm' without measuring F0'. Photosynthesis Research 54: 135-142.

NPQ Protocol




Fig. 9 NPQ Protocol.

5.3.3 LIGHT CURVE (LC) PROTOCOLS

The protocols called Light Curve (LC) were designed to acquire parameters for construction of Light Response Curve relating the rate of photosynthesis to photon flux density. The method is based on successive measurements of the sample exposed to a stepwise increase of light intensity. The effective quantum yields of photosynthesis are determined under various light intensities of continuous illumination. Measurement is based on pulse modulated fluorometry (PAM).

Three predetermined LC protocols are available. These differ in number and duration of individual light phases and light intensities as shown in Table 3 below. The visual representation of the LC1 and LC2 protocols is shown in Fig. 10 and Fig. 11 below.

	# of phases	Phase duration	Light intensities [$\mu\text{mol}\cdot\text{m}^{-2}\cdot\text{s}^{-1}$]
LC1	6	60s	10; 20; 50; 100; 300; 500
LC2	5	30s	100; 200; 300; 500; 1000
LC3	7	60s	10; 20; 50; 100; 300; 500; 1000

Table 3 LC protocols.

The LC protocols include the following measured and calculated parameters:

Abbreviation	Explanation
F_0	minimum fluorescence in dark-adapted state
F_m	maximum fluorescence in dark-adapted state
$F_{m_Ln}^\ddagger$	maximum fluorescence in light adaptation state
$F_{t_Ln}^\ddagger$	instantaneous fluorescence during light adaptation
QY_{max}^*	maximum quantum yield of PSII in dark-adapted state - F_v/F_m
QY_{Ln}^{***}	instantaneous PSII quantum yield induced in light

\ddagger n represents a sequential number of light phases

*Calculated as $(F_m - F_0) / F_m$

** Calculated as $(F_{m_Lx} - F_{t_Lx}) / F_{m_Lx}$

Light Curve 1 Protocol




Fig. 10 LC1 Protocol.

Light Curve 2 Protocol




Fig. 11 LC2 Protocol.

5.4 MULTIPLE MEASUREMENT

In addition to the single measurement of each available protocol, it is possible to perform also the multiple measurements of the same protocol over a period of time. The FluorPen may be set up to perform repeated measurements of the same parameter or protocol by selecting **Settings > Multi** (see Menu tree, page 27):

Multi type – select the required parameter or protocol - Ft, QY, OJIP...

Multi interval – set the time interval between measurements (the interval represents period from the protocol beginning to the next protocol beginning and so, the interval should be longer than the duration of the selected protocol)

Multi repeats – set the number of measurements

- Prepare the sample for measurement as described above.
- Select in the display menu: **Measurement > Multi**.
- Press **SET** to start the measurements.
- Values of measured parameters (Ft, QY) will appear on the FluorPen display after each measurement repetition and will be stored to the device memory automatically. If protocol (OJIP, NPQ, LC) was used all data will be saved to the device memory and visualization will be possible after the data download to the PC (page 44).

Modes of Multiple measurement:

There are two modes of the multiple measurements in the FluorPen.

1. The device is connected via USB to the computer.

The device measures according to the predefined protocol, interval and repeats. The device does not switch off between measurements and display a progress on the front control display. After reaching of the predefined number of repeats, the device turns off the Multiple measurement automatically.

2. FluorPen is not connected to the computer.

The device measures according to the predefined protocol and interval but it doesn't stop the Multiple measurement after reaching of the predefined repetitions: the device measures continuously as long as it is stopped. Also, the device turns on and off between the measurements automatically.

The Multiple measurement can be stopped manually by switching on the device (via the SET button) during the switched off period. Should the device be in the measuring mode, then the measurement needs to be interrupted via a long press of the MENU button and after that the Multiple measurement needs to be stopped by switching on the device via the SET button.


6 CONTROL MENU TREE

The next few pages of this manual show the structure of the menu and explain in a schematic way the operation of the FluorPen. The schematic diagrams show the Main Menu, first-level Sub-Menus and second-level Sub-Menus.

- The blue color represents the Main Menu and its Options.
- The yellow color represents the first-level Sub-Menus and their Options.
- The green color represents the second-level Sub-Menus and their Options.
- Full-line arrows are used to indicate the **SET** key operations.
- Dashed-line arrows are used to indicate the **MENU** key operations.


Main Menu

To start hold the SET key for 1 second.




Measure Sub-Menu - Part 1

Use the Measure Sub-Menu when measuring selected parameters.




Measure Sub-Menu - Part 2

Use the Measure Sub-Menu when measuring selected parameters.



Data Sub-Menu


Use the Data Sub-Menu when browsing or erasing previously measured data.



IMPORTANT NOTE: Be aware that it is not possible to erase single data.
All stored data are erased!


Setting Sub-Menu - Part 1

Use the Setting Sub-Menu to set the light color, light intensity, number and frequency of measurements, date, time, or the sound mode.



Setting Sub-Menu - Part 2


Use the Setting Sub-Menu to set the light color, light intensity, number and frequency of measurements, date, time, or the sound mode.



To be continued on the next page

Setting Sub-Menu - Part 3

Use the Setting Sub-Menu to set the light color, light intensity, number and frequency of measurements, date, time, or the sound mode.



7 USB CONNECTION

FluoPen comes with the USB cable that is required for charging of the Li-ion battery and can also be used for data transfer to the PC after completion of measurements. To connect the USB cable with the FluoPen device follow the picture instructions below. Please note that a lock in system is used to secure the USB cable to the FluoPen and extreme caution has to be used when setting up this connection to avoid damage to the cable pins.



When connecting the USB cable take extra caution to prevent damage to the cable connector pins. Ensure correct orientation of the cable as shown in the pictures below so the circled portion of the plug and the cable in photo A and B are perfectly lined up prior to pushing them together. Once this connection is achieved the cable may be secured in position by turning the metal cover of the cable and locking the cable in position.

To connect FluoPen with your computer please follow steps below in

Fig. 12:




Fig. 12 How to connect FluoPen with PC.

A) connector on the FluoPen device. B) Portion of the USB cable with pins. C – E) Position the cable horizontally and line up the green circled parts of the cable and the connector, plug in the inlet and screw the securing screw. F) Correct connection of the USB cable and Pen device.

Once the cable is securely attached to the FluoPen the other end may be connected to the USB port on a PC. The FluoPen **switches ON** automatically after connecting the cable to the PC. For the USB connection to be successful the USB driver and the FluoPen software need to be installed on the PC. Both may be found on the installation disk (USB driver folder) delivered with the device. Once the USB driver is installed the Device Manager in Windows will list the USB serial port in the device tree. The USB driver may also be downloaded from PSI websites www.psi.cz. Once the driver is installed correctly the connection between the FluoPen and the computer is initiated by selecting in the software on the computer **Setup > Device ID**.

For more information about FluoPen software see chapter 9.



Monitoring Pen device equipped with Battery Pack has to be connected to PC through Battery Pack.

8 BLUETOOTH CONNECTION

In addition to data transfer via USB the FluorPen may be connected to the software via Bluetooth for data transfer. Before setting up the Bluetooth connection between the FluorPen and the PC, ensure the following components are in place:

Bluetooth enabled PC

The PC must have Bluetooth wireless technology, either built-in or through a Bluetooth card. Ensure that the PC's Bluetooth setting is in "discoverable" mode (meaning that it shows up when other devices search for nearby Bluetooth connections). Consult the user guide for the PC or Bluetooth card to learn how to do this.

Bluetooth configuration software properly set up on the PC

Before you connecting the FluorPen to the PC and downloading data files the Bluetooth software that came with the PC, or the PC's Bluetooth card needs to be activated. This software varies by manufacturer. Please consult the PC's Bluetooth documentation for more information.

Bluetooth must be switched on and be visible on both devices

To pair the FluorPen with another Bluetooth device, such as a computer, ensure that Bluetooth is switched on visible on both devices.

8.1 BLUETOOTH PAIRING

1. Enabling Bluetooth on the FluorPen

- Switch ON the FluorPen (press and hold the SET key for 1 s).
- Scroll to the Accessories menu (press the MENU key) and select Accessories by pressing the SET key.
- Select Bluetooth On (press the MENU key, then turn it ON by pressings the SET key).



Keep in mind that the FluorPen turns off automatically after about 8 minutes of no action.
Turning off the FluorPen always turns Bluetooth off.

2. Starting Bluetooth Application on the PC

- The following description of how to set up the Bluetooth connection between the PC and the device is for Windows 7; some of the steps may be different with different version of Windows.
- Select: Start > Devices and Printers (Fig. 13).
- You may also start your Bluetooth application via the Control Panel: Start > Control Panel > Hardware and Sound > Devices and Printers.




Fig. 13 Start bluetooth application.

3. Opening the Add Bluetooth Device Application

- Select: “Add a device” to start searching for the new Bluetooth device. Be sure that the FluorPen is in discoverable mode (see step 1).



Fig. 14 Adda device.

4. Selecting the FluorPen

- Select: PSI FluorPen icon.
- Click: Next (Fig. 15).




Fig. 15 Select the FluorPen.

5. Starting the Pairing Process

This step is different for old and new version of the FluorPen, that are equipped with disparate Bluetooth module.

Old version of FluorPen (FP-100):

Your Bluetooth Pairing Code is: 0000

- Select: “Enter the device’s pairing code”.
- Enter: 0000 (four digits).
- Select: Next (Fig. 16).




Fig. 16 Pairing process.

New version of FluorPen FP-110:

- Select: Yes (Fig. 17). Please note that the FluorPen device does not display the verification number. The verification code is not important for the BT connection.
- Select: Next.




Fig. 17 Verifying of the Bluetooth pairing.

6. Completing the FluorPen Pairing

- Select: Close (Fig. 18).




Fig. 18 Finishing.

The Bluetooth pairing is now complete, and the next step is to open the program FluorPen 1.1 (included on the USB flash disk) For more information about FluorPen software see chapter 9.

9 FLUORPEN SOFTWARE

9.1 SOFTWARE INSTALLATION

1. Copy the FluorPen software provided on the USB flash disk to your computer and launch the FluorPen program.
2. To connect and recognize the FluorPen device in the FluorPen software, proceeds first with the registration of the FluorPen software (Fig. 19).
 - Select: Help > **Register**
 - Enter: your serial registration number (found in a text file on the USB flash disk drive included with the device).
 - Select: OK




Fig. 19 Software registration.



Please note that the serial (registration) number for the FluorPen may be found in the file **SN.txt**, which is included on the enclosed USB flash disk.

Please Note: it is not possible to download data from the FluorPen device without software registration.

3. Switch on the FluorPen and enable Bluetooth or connect USB cable to the PC.
4. Ensure the PC and the FluorPen are properly paired (see chapter 7 and 8 for complete information on USB and Bluetooth pairing).
5. In the software select: **Setup > Device ID (Ctrl+I)**. If properly connected, the message "Device: FluorPen" appears in the bottom part of the screen (Fig. 20). If the connection is not successful then message "Device not found" will appear. In the latter case check all the connections (USB) and Bluetooth pairing.




Fig. 20 Connecting FluorPen with software.

9.2 MENU AND ICONS EXPLANATION

9.2.1 MAIN MENU

MENU: File

Load	Loads previously saved data files.
Save	Saves data to hard disc.
Export	Exports data in .txt format.
Export to JSON	Exports data in JavaScript Object Notation.
Close	Closes the current experiment.
Close All	Closes all running experiments.
Exit	Exits the program.



MENU: Device

Download	Downloads data from the FluorPen to your PC.
Erase Memory	Erases data from the FluorPen memory.
Online Control	Online control of FP device.
Attach GPS File	Used for download data from GPS module (active only in FluorPen version FP 100 and Monitoring Pen MP 100).




MENU: Setup

Device ID	Detects the connected device.
Update Firmware	Used for firmware updates.
Settings	Used for modification of the program settings.



MENU: Help

About	Offers basic information about the program.
Register	Used for the FluorPen software registration.



Icon Explanation:



Download	Downloads data from the FluorPen to PC.
Load	Loads (opens) previously saved data files.
Save	Saves data to hard disc.
Export	Exports data in .txt format.

9.2.2 MENU SETTINGS

MENU > Setup > Settings (Fig. 21)

After Download – Memory Erase

If the box is checked the FluorPen memory is erased after each data download.

Data – Inverted

If the box is checked the polarity of data is inverted, e.g., multiplied by -1. This feature can be helpful for a certain type of experiment when the measured data are undesirably interpreted as negative values.

Data – Add to opened

If the box is checked the downloaded data are added to that of the current opened experiment.

Data – GPS Visible

This option is active only in older FluorPen version FP 100 and Monitoring Pen MP 100. In new version FP 110 the GPS data are automatically downloaded and paired with protocol measurements.

Graf – Single

If the box is checked all measured data are visualized in one graph, i.e., the value of each new measurement is added to the currently used graph window.

If the box is not checked a new graph is opened for every new measurement.




Fig. 21 Settings.

9.2.3 MENU ONLINE CONTROL

This function can be used for Online Control the FluorPen device after connection with the PC.

- Select: **Menu > Device > Online Control**

Online Control – Switches (Fig. 22)

Sound On/Off – select presence of sound - device beeping when pressing MENU and SET keys.

Multi use averaging (YES/NO) – serves to confirm **Multi Repeats** (number of repetitions) and **Multi Interval** (time between measurements). Settings for each measurement within **Multi** Option preset by the user on the FluorPen device or in the software under Values tab (see below)– select YES or NO.



Fig. 22 Online control - Switches.

Online Control – Values (Fig. 23)

The tab Values allows to set the measuring characteristics as light pulses intensities, measurement averaging and Multiple measurement. **Please remember that each set value needs to be confirmed either by leaving the box or by pressing the ENTER key.**

The intensities of Actinic, Super and Flash Pulse light can be set in this window.

Averaging enables to measure specific consecutive protocols and to **calculate arithmetic mean of the measurements**. The measurement result is represented by one average value or graph. The number of the measurement (**Averag. Count**) and the delay between the individual measurements (**Averag. delay**) need to be set. The averaging is not available for the QY measurement. Please remember, that this function is not suitable for the dark-adapted samples (excepting the Ft measurement) as the sample is illuminated by the appropriate light pulses during the first measurement.

Multiple measurement is automatically repeated measurement of one specific protocol over a time period. The measurement result is represented by the values or graphs obtained in the specific data points. The **Multi interval** represents the time between the beginnings of the single measurements and so, the Multi interval should be longer than the duration of the selected protocol. The number of **Multi repeats** from 1-1000 can be set in this window. Finally, the type of Protocol selected for Multiple measurements (**Multi type**) is also set in this window (see Fig. 23). Please note that the Multi measurements have to be started from the device or by clicking on the **Multi** button in the “Protocols” tab of the Online Control window (see Fig. 25).




Fig. 23 Online control – Values.

Online Control – Time (Fig. 24)

The FluorPen time and date can be set in this window. The time and date can be edited either manually by setting and saving it or synchronized with the computer automatically by ticking the box. The synchronization is performed just at once (i.e., the option doesn't synchronize the time continuously). This is essential for correct GPS data acquisition and therefore recommended.




Fig. 24 Online control - Time.

Online Control – Protocols (Fig. 25)

Selection of the protocol for single measurements may be done under this tab in the software. Once the measurement is completed the data is saved to the device and can be downloaded to the PC at a later time. Measuring of PAR is active only in the PAR-FluorPen FP 110 (see picture below). Also, by pressing the Multi button in this window the Multiple measurements can be started remotely.




Fig. 25 Online control - Protocols.

9.3 DATA TRANSFER AND VISUALIZATION

1. Once kinetic protocols data (OJIP, NPQ, LC) have been collected with the FluorPen to visualize the data it needs to be downloaded to the PC first via FluorPen software. Before data transfer can occur a successful connection between the FluorPen and the PC needs to be established via USB cable or Bluetooth module (see chapter 7 and 8 for details).
2. Click the **Download** icon or select **Device > Download**.
3. Once the download is complete the Data can be visualized in a table shown below (Fig. 26).




Fig. 26 Example of data transfer and visualization.

4. To visualize the data in the graph mode, click the **Graph** field in the bottom bar.
5. The selected set of data will be shown on the Graph (Fig. 27).




Fig. 27 Graphic visualization of experiment.

6. To export data from the FluorPen software select **File > Export** or **Export** icon. Select data type to export (Ft, QY, OJIP...) - Fig. 28.

Selected only – exports only one measurement that is selected by mouse, otherwise it will export everything.

Source data – exports raw data, in case of OJIP: points of the curve.

Description – exports the data description if any.

Computed values – export calculated data, in case of OJIP: F_0 , F_i , F_j ...

Index	4	5	6	7	8	9
Time	19:01:05 18.4.2018	19:33:50 18.4.2018	20:01:00 18.4.2018	20:33:51 18.4.2018	21:01:02 18.4.2018	21:33:49 18.4.2018
	OJIP-455	NPQ1-455	OJIP-455	NPQ1-455	OJIP-455	NPQ1-455
Bdkg	25813	Fo 2634	Bdkg 290	Fo 2698	Bdkg 195	Fo 2698
Fo	6795	Fm 7215	Fo 6469	Fm 7640	Fo 6632	Fm 7673
Fj	10794	Fp 6242	Fj 11736	Fp 7087	Fj 12159	Fp 7185
Fi	13687	Fm_L1 4909	Fi 15670	Fm_L1 5137	Fi 15832	Fm_L1 5202
Fm	14825	Fm_L2 4259	Fm 17653	Fm_L2 4617	Fm 17913	Fm_L2 4714
Fv	8030	Fm_L3 4031	Fv 11184	Fm_L3 4259	Fv 11281	Fm_L3 4422
Vj	0.498				Vj 0.490	Fm_L4 4292
Vi	0.858				Vi 0.816	Fm_Lss 4227
Fm/Fo	2.182				Fm/Fo 2.701	NPQ_L1 0.48
Fv/Fo	1.182				Fv/Fo 1.701	NPQ_L2 0.63
Fv/Fm	0.542				Fv/Fm 0.630	NPQ_L3 0.74
Mo	1.166				Mo 1.051	NPQ_L4 0.79
Area	1560378				Area 2551333	NPQ_Lss 0.82
Fix Area	14353626				Fix Area 17536264	Qp_L1 0.15
Sm	194.319				Sm 226.162	Qp_L2 0.17
Se	0.427				Se 0.462	Qp_L3 0.20
N	455.013				N 489.561	Qp_L4 0.22
Phi_Po	0.542				Phi_Po 0.630	Qp_Lss 0.32
Phi_o	0.502				Phi_o 0.510	Rfd 0.91
Phi_Eo	0.272				Phi_Eo 0.321	Fm_D1 7055
Phi_Do	0.458				Phi_Do 0.370	Fm_D2 7510
Phi_Pav	942.951				Phi_Pav 939.376	Fm_D3 7543
Pl_Abs	0.276				Pl_Abs 0.515	NPQ_D1 0.09
ABS/Rc	4.323				ABS/Rc 3.437	NPQ_D2 0.02
TRo/Rc	2.342				TRo/Rc 2.165	NPQ_D3 0.02
ETo/Rc	1.175				ETo/Rc 1.104	Qp_D1 0.92
...						...

Fig. 28 Export data window.

9.4 FIRMWARE UPDATE



All data in the FluorPen memory are erased during the firmware update!

Before starting any firmware update, download all your data from the FluorPen memory to the computer!

1. Starting Update

- Select: **Setup > Update Firmware From File** (Fig. 29).




Fig. 29 Firmware Update.

2. Warning

- Select: **OK** to start update (Fig. 30)




Fig. 30 Warning.

3. Selecting .bxn file

- Find: firmware update file: Binary file (with the extension .bxn) (Fig. 31).
- Select: Open.




Fig. 31 Select .bxn file.

4. Finishing Upload

- Select: **OK** to start uploading of the update (Fig. 32).




Fig. 32 Data loss warning.

- The bottom bar indicates the upload progress (Fig. 33).



Fig. 33 Upload progress.

- Press: **OK** to finish upload (Fig. 34).





Fig. 34 Finish upload.

10 GPS MODULE

The new versions of the FluorPen device (FP110) have integrated GPS module which can be turned on during the measurements. When GPS module is turned on the map coordinates will be automatically saved with all collected data and will be downloaded during data download.

	For proper GPS reading, the time in your FluorPen and in your computer must be synchronized. Preset time and time zone must correspond to GPS time (time zone) in your location.
---	--


10.1 GPS/FLUORPEN OPERATION


1. Check the time setting on the FluorPen device: **Settings > Date & Time**
2. Switch the GPS module "ON" on the FluorPen device by following these steps in the menu:
 - Select: **Accessories > GPS**
 - Press **SET** to confirm.
 - Wait until the GPS position is found – "**Starting GPS**".
 - The GPS module is ready when the icon in upper panel changes as shown on Fig. 35.



Fig. 35 GPS icons.

3. If the picture on the display of the device does not change then proceed to **Accessories>GPS>Location** selection in the menu and manually map the GPS by pressing SET. "**GPS Acquisition**" message will appear followed by coordinate. If the GPS module has difficulties mapping the coordinates, a message stating "**GPS not locked**" will appear on the display. It may be necessary to take the device outside into a location that is easily accessible by the satellite (clear sky view) and repeat the process of mapping.
4. Once the GPS has been turned on and successfully activated proceed to **Measurement** and select required protocol.

	For prompt determination of the coordinates use the option Accessories > GPS > Location .
---	--

	The device may need a clear view of the sky to acquire satellite signal. Keep in mind that the FluorPen turns off automatically after about 8 minutes of no action. Turning off the FluorPen always turns off GPS module.
---	---

10.2 DATA DOWNLOAD

1. Enabling Communication:
 - Switch on the computer and the FluorPen. Set your computer to FluorPen communication: enable Bluetooth or connect to USB port (see instructions on pg. 35).
2. Downloading Data from the FluorPen
 - Start FluorPen program.
 - Connect FluorPen device: **Setup > Device ID (Ctrl+I)**
 - Download measured data from the FluorPen to your PC by clicking the down lad icon (top icon). Data measured with activated GPS module are downloaded with GPS coordinates (Fig. 36).

Index	2	3	4	5	6	7
Time	10:27:54 29.3.2018	10:29:29 29.3.2018	10:31:45 29.3.2018	10:35:52 29.3.2018	10:22:44 3.4.2018	10:23:11 3.4.2018
	49° 20.3871' N 16° 28.6379' E	49° 20.3538' N 16° 28.6755' E	49° 20.2923' N 16° 28.6290' E	49° 20.2557' N 16° 28.5246' E	Qy 0.67	Qy 0.04
Value	Qy 0.72	Qy 0.65	Qy 0.27	Qy 0.67	Fo Backgr 378 Fo Flash 3310	Fo Backgr 897 Fo Flash 976
	Fo Backgr 299 Fo Flash 4985	Fo Backgr 378 Fo Flash 2711	Fo Backgr 89 Fo Flash 1069	Fo Backgr 438 Fo Flash 3110	Fm Backgr 398 Fm Flash 9331	Fm Backgr 864 Fm Flash 946
	Fm Backgr 299 Fm Flash 17138	Fm Backgr 418 Fm Flash 7058	Fm Backgr 92 Fm Flash 1436	Fm Backgr 418 Fm Flash 8544		
Description						

Fig. 36 GPS coordinates.

11 WARRANTY TERMS AND CONDITIONS

- This Limited Warranty applies only to the FluorPen device. It is valid for one year from the date of shipment.
- If at any time within this warranty period the instrument does not function as warranted, return it and the manufacturer will repair or replace it at no charge. The customer is responsible for shipping and insurance charges (for the full product value) to PSI. The manufacturer is responsible for shipping and insurance on return of the instrument to the customer.
- No warranty will apply to any instrument that has been (i) modified, altered, or repaired by persons unauthorized by the manufacturer; (ii) subjected to misuse, negligence, or accident; (iii) connected, installed, adjusted, or used otherwise than in accordance with the instructions supplied by the manufacturer.
- The warranty is return-to-base only and does not include on-site repair charges such as labor, travel, or other expenses associated with the repair or installation of replacement parts at the customer's site.
- The manufacturer repairs or replaces faulty instruments as quickly as possible; the maximum time is one month.
- The manufacturer will keep spare parts or their adequate substitutes for a period of at least five years.
- Returned instruments must be packaged sufficiently so as not to assume any transit damage. If damage is caused due to insufficient packaging, the instrument will be treated as an out-of-warranty repair and charged as such.
- PSI also offers out-of-warranty repairs. These are usually returned to the customer on a cash-on-delivery basis.
- *Wear & Tear Items* (such as sealing, tubing, padding, etc.) are excluded from this warranty. The term *Wear & Tear* denotes the damage that naturally and inevitably occurs as a result of normal use or aging even when an item is used competently and with care and proper maintenance.

12 TROUBLESHOOTING AND CUSTOMER SUPPORT

In case of problems with the FluorPen visit [FAQ](#) on our websites (<http://psi.cz/support/faq>) or contact customer support by email to support@psi.cz, or contact your local distributor.

13 APPENDIX

13.1 BATTERY PACK FOR MONITORING PEN

Battery pack serves as an external power source for Monitoring Pen device. The external battery provides power during the long-term experiments.



Please note that Monitoring Pen devices equipped with battery pack do not have internal battery, therefore it is not possible to use them without the battery pack.

13.1.1 STANDARD BATTERY PACK

Standard battery pack (Fig. 37) is intended for the operation within temperature range from +10 °C to +40 °C. The operating time is up to 2 years (QY measurement every 1 hour). The pack includes battery case with rechargeable sealed lead acid battery (12Ah), charger, two types of cables (serial and device) and serial convertor.



Fig. 37 Standard battery pack.

Connectors of Standard battery pack (Fig. 38):

FluorPen - connects the Monitoring Pen to the battery. This connection is necessary for Monitoring Pen operation and data download, it provides power to the Monitoring Pen.

Serial - enables communication between Monitoring Pen and PC for control and data download.

Charger – enables to charge the battery.



Fig. 38 Connectors of Standard battery pack.

Replacement of the battery:

If the battery needs to be changed follow these steps:

1. Unscrew 4 screws (in each corner of the battery casing) and remove the battery pack cover (Fig. 39 a).
2. Disconnect the internal battery from the cables.
3. Place new battery inside the casing, connect it with the cables – red cable with red and black cable with black marked connector (Fig. 39 b, c). Replace the cover.




Fig. 39 Replacement of battery in a Standard Battery Pack.

13.1.2 EXTENDED TEMPERATURE RANGE BATTERY PACK

Extended temperature range battery pack (Fig. 40) is intended for operation within temperature range from $-40\text{ }^{\circ}\text{C}$ to $+60\text{ }^{\circ}\text{C}$. Operating time is up to 2 years (QY measurement every 1 hour). The pack includes battery case with non-rechargeable Li-SOCl₂ battery (5.5Ah), two types of cables (serial and device) and serial convertor.



Fig. 40 Monitoring Pen MP 100-A charged via Extended temperature range battery pack.



Extended temperature range battery pack cannot be recharged. Spare battery is offered as additional accessory.

Connectors of Extended temperature range battery pack:

Device - provides power connection between Monitoring Pen and the battery. This connection is necessary for Pen operation and data download.

Serial - enables communication between Monitoring Pen and the PC for control and data transfer.

Replacement of the battery:

Follow these steps to replace the battery:

1. Remove 4 screws in the corner of the case and remove the battery pack cover (Fig. 41 a, b).
2. Disconnect the internal battery from the cable (Fig. 41 c).
3. Place the new battery inside the casing, connect it with the cable and replace the cover.




Fig. 41 Extended temperature range battery pack: battery replacement.

13.2 INSTALLATION AND OPERATION OF THE MONITORING PEN MP 100-A/B

13.2.1 DEVICE CONTROL

To **turn on** the device connect the serial cable to the battery pack and pc.

The device **stays on** when the serial cable is connected.

The Monitoring Pen automatically **turn off** after serial cable disconnection.

Protocol setting is possible only using Online control in the FluorPen software.

Multiprotocol serves for automated measurement of the predefined protocols in the predefined time interval. After setting and starting of the Multiprotocol, the serial cable can be disconnected. The device switches off during the measurements automatically in order to save the battery. The Multiprotocol measures continuously (regardless of the set Multi repeats) until the battery is discharged or the protocol is stopped manually. More information about the Multiprotocol can be found in chapter 9.2.3.

Online control enables checking of **battery status** (Fig. 42). The current battery voltage is shown as measured data in FluorPen software (Fig. 43).




Fig. 42 Online control window enables checking of the battery status.

Index	2	3	4	5	6
Time	11:01:20 24.8.2018	11:01:42 24.8.2018	11:01:46 24.8.2018	11:01:54 24.8.2018	11:02:02 24.8.2018
	Ft 199	Voltage [V] 6.25	Voltage [V] 6.25	Voltage [V] 6.25	Voltage [V] 6.25
Value	Backgr 14486 Flash 14685				
Description					

Fig. 43 Battery status.



The minimal battery voltage is 4.9 V. The device does not work at lower voltage. Please find more information about the battery pack in chapter 13.1.

13.2.2 DEVICE INSTALLATION

1. Connect the Monitoring Pen to the cable.
2. Place the Monitoring Pen under the water.
3. Connect the device cable to the battery pack.
4. For single measurement and multiprotocol setting connect the serial cable to the battery pack and pc.
5. Open the FluorPen software and connect the device (more details in chapter 9).
6. Set the protocol using the Online control and start the protocol.
7. Disconnect the serial cable. The device switches automatically to standby mode and measures according to preset protocol.
8. For data download connect the serial cable. Connection of serial cable cancels the multiprotocol measurement. Start the multiprotocol again for following measurement.



Please note that only the Monitoring Pen and device cable are submersible.

13.3 DETACHABLE LEAF CLIPS

Detachable leaf clips are used with the FluoPen FP 110/D and PAR-FluorPen FP110/D for dark adaptation of the leaf before measurements of chlorophyll fluorescence. Multiple leaf clips may be placed on leaves in a closed position ahead of time to allow dark adaptation while measurements of other leaves proceed. Start with the clip in a closed position (metal screen covering the leaf). Attach the FluorPen probe to the clip by pressing it into the clip. Once the connection is secure, slide the metal screen to expose the leaf to the FluorPen's optical probe. Proceed with the measurements. See Fig. 44 for visual of the leaf clip in a closed and open position. Detachable leaf clips may be purchased in sets of 10.




Fig. 44 The detachable leaf clip in open (A) and closed (B) position.